
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002484
8:00 21/10/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:03
136.0 C 220kPa 3:03
136.3 C 223kPa 3:33
136.0 C 220kPa 4:03
136.2 C 221kPa 4:33
135.8 C 220kPa 5:03
136.3 C 222kPa 5:33
136.2 C 220kPa 6:03
136.4 C 221kPa 6:33
136.4 C 221kPa 6:33
VENTING 6:33
AIR DRYING 7:23
CYCLE COMPLETE 8:34

Digital Signature #
94AFEE346DD388ED
