

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3217

Name: RITTER 1  
Sterilizer ID: V2178198  
05/11/2023 08:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.3 F 0.0  
2:00 150.8 F 0.0  
4:00 160.3 F 0.9  
6:00 194.9 F 5.2  
8:00 212.2 F 10.2  
10:00 223.5 F 6.3  
12:00 233.2 F 9.5  
14:00 238.9 F 9.7  
16:00 242.0 F 12.0  
18:00 240.1 F 10.4  
20:00 253.6 F 17.2  
22:00 265.4 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.7 F 29.1  
2:00 271.5 F 28.4  
3:00 272.4 F 28.8  
4:00 271.7 F 28.6

Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 231.2 F 0.0  
10:00 231.7 F 0.0  
15:00 227.6 F 0.0  
20:00 225.5 F 0.0  
25:00 223.9 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:38  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:01:55  
CYCLE COMPLETE

05/11/2023 09:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_