
Midmark M11 - v1.0.5
Total Cycles: 3575

Name: RITTER2
Sterilizer ID: V2313228
05/01/2024 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.3 F 0.0
2:00 154.7 F 1.9
4:00 157.0 F 5.1
6:00 180.1 F 8.8
8:00 199.6 F 13.0
10:00 212.2 F 6.4
12:00 225.5 F 10.6
14:00 229.4 F 7.9
16:00 238.4 F 12.6
18:00 239.3 F 11.0
20:00 251.3 F 17.0
22:00 261.4 F 23.2
24:00 271.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.9
1:00 272.3 F 30.5
2:00 272.9 F 30.8
3:00 272.9 F 30.6
4:00 272.8 F 30.3

Min 271.4 F 29.7
Max 273.3 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 219.0 F 0.0
10:00 232.1 F 0.0
15:00 223.3 F 0.0
20:00 209.9 F 0.0
25:00 206.9 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:47
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:21
CYCLE COMPLETE

05/01/2024 12:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____