
Midmark M11 - v1.0.5
Total Cycles: 3953

Name: RITTER2
Sterilizer ID: V2313228
09/18/2024 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.1 F 0.0
2:00 133.9 F 1.5
4:00 143.7 F 4.0
6:00 170.2 F 7.8
8:00 204.1 F 12.1
10:00 214.4 F 5.7
12:00 228.4 F 10.7
14:00 231.4 F 7.5
16:00 241.0 F 13.1
18:00 241.8 F 11.9
20:00 255.2 F 18.4
22:00 266.2 F 25.2
24:00 271.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.7
1:00 272.0 F 29.8
2:00 272.1 F 29.9
3:00 272.0 F 30.0
4:00 272.0 F 29.9

Min 271.8 F 29.5
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.1 F 0.0
5:00 246.9 F 0.0
10:00 232.7 F 0.0
15:00 228.2 F 0.0
20:00 227.7 F 0.0
25:00 229.8 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:07
CYCLE COMPLETE

09/18/2024 03:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____