

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4058

Name: RITTER 1  
Sterilizer ID: V2178198  
11/27/2023 10:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.7 F 0.0  
2:00 131.3 F 0.9  
4:00 129.5 F 3.0  
6:00 153.5 F 6.0  
8:00 175.1 F 9.4  
10:00 203.5 F 13.2  
12:00 212.0 F 5.2  
14:00 223.9 F 9.1  
16:00 232.4 F 13.5  
18:00 235.6 F 9.7  
20:00 238.3 F 9.9  
22:00 244.8 F 13.0  
24:00 256.0 F 18.8  
26:00 266.1 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.2  
2:00 271.3 F 28.5  
3:00 272.4 F 29.3  
4:00 271.7 F 28.7

Min 271.2 F 28.5  
Max 272.4 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 209.5 F 0.0  
10:00 210.4 F 0.0  
15:00 210.4 F 0.0  
20:00 217.7 F 0.0  
25:00 216.8 F 0.0  
30:00 219.0 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 27:47  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:06:24  
CYCLE COMPLETE

11/27/2023 12:02 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_