

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2430

Name: RITTER2  
Sterilizer ID: V2313228  
05/29/2023 02:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
2:00 198.9 F 3.9  
4:00 209.5 F 9.2  
6:00 226.0 F 8.2  
8:00 233.6 F 11.4  
10:00 235.2 F 9.2  
12:00 239.4 F 9.5  
14:00 248.6 F 14.9  
16:00 260.9 F 21.9  
18:00 271.9 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.0  
1:00 272.3 F 29.8  
2:00 271.7 F 29.1  
3:00 271.3 F 29.3  
4:00 272.5 F 29.8  
  
Min 271.0 F 28.8  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 221.9 F 0.0  
10:00 242.1 F 0.0  
15:00 232.1 F 0.0  
20:00 227.5 F 0.0  
25:00 226.1 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:58  
CYCLE COMPLETE

05/29/2023 02:58 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_