

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1273

Name: RITTER 1  
Sterilizer ID: V2178198  
10/18/2021 10:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
2:00 182.4 F 2.7  
4:00 185.1 F 6.9  
6:00 210.4 F 12.9  
8:00 222.7 F 8.1  
10:00 236.6 F 12.2  
12:00 240.3 F 11.1  
14:00 237.7 F 9.3  
16:00 251.7 F 16.2  
18:00 263.1 F 23.1  
20:00 272.5 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.9 F 29.8  
2:00 272.4 F 29.4  
3:00 271.8 F 28.8  
4:00 271.6 F 29.1

Min 271.3 F 28.7  
Max 273.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 238.2 F 0.0  
10:00 221.3 F 0.0  
15:00 208.7 F 0.0  
20:00 207.0 F 0.0  
25:00 202.5 F 0.0  
Dry Cycle Aborted

FILLING: 1:21  
HEATING: 20:18  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 25:11  
TOTAL CYCLE: 00:53:54

10/18/2021 11:02 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01