
Midmark M11 - v1.0.0
Total Cycles: 5169

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2024 12:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 147.7 F 0.1
4:00 163.3 F 0.4
6:00 185.7 F 1.2
8:00 206.1 F 4.4
10:00 221.2 F 8.5
12:00 232.7 F 12.8
14:00 234.1 F 8.5
16:00 244.0 F 12.6
18:00 243.2 F 11.8
20:00 240.7 F 10.5
22:00 253.4 F 16.8
24:00 263.7 F 22.8
26:00 272.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.0 F 28.5
2:00 272.7 F 29.0
3:00 272.6 F 28.8
4:00 272.2 F 28.5

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 227.5 F 0.0
10:00 226.1 F 0.0
15:00 189.1 F 0.0
20:00 180.3 F 0.0
25:00 178.0 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:26
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:06
CYCLE COMPLETE

12/06/2024 01:55 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____