
Midmark M11 - v1.0.5
Total Cycles: 1694

Name: RITTER2
Sterilizer ID: V2313228
11/07/2022 08:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.2 F 0.0
2:00 197.1 F 3.1
4:00 212.0 F 9.4
6:00 223.9 F 7.0
8:00 237.6 F 11.8
10:00 238.7 F 10.4
12:00 238.6 F 9.8
14:00 254.1 F 17.4
16:00 266.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.7
2:00 272.6 F 29.3
3:00 271.8 F 28.4
4:00 272.8 F 29.2

Min 271.5 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 221.4 F 0.0
10:00 229.0 F 0.0
15:00 229.2 F 0.0
20:00 223.9 F 0.0
25:00 221.4 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 17:24
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:56:53
CYCLE COMPLETE

11/07/2022 08:59 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____