

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2256

Name: RITTER2  
Sterilizer ID: V2313228  
04/14/2023 03:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.1 F 0.0  
2:00 149.1 F 2.1  
4:00 156.7 F 5.1  
6:00 179.4 F 9.0  
8:00 199.2 F 13.5  
10:00 229.2 F 7.3  
12:00 233.7 F 11.9  
14:00 235.4 F 8.7  
16:00 242.4 F 11.4  
18:00 244.9 F 12.8  
20:00 256.1 F 18.8  
22:00 266.4 F 25.1  
24:00 272.5 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 271.5 F 29.4  
2:00 271.5 F 29.4  
3:00 271.7 F 29.3  
4:00 271.8 F 29.4

Min 271.4 F 29.3  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
5:00 228.5 F 0.0  
10:00 224.8 F 0.0  
15:00 225.9 F 0.0  
20:00 224.8 F 0.0  
25:00 226.9 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:02:53  
CYCLE COMPLETE

04/14/2023 04:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_