
Midmark M11 - v1.0.5
Total Cycles: 3115

Name: RITTER2
Sterilizer ID: V2313228
11/13/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.5 F 0.0
2:00 187.0 F 1.5
4:00 187.5 F 5.7
6:00 202.7 F 10.6
8:00 219.0 F 6.0
10:00 226.0 F 10.4
12:00 230.2 F 7.8
14:00 242.4 F 13.3
16:00 242.7 F 12.1
18:00 255.0 F 18.5
20:00 266.0 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.6
2:00 271.4 F 28.9
3:00 272.5 F 29.8
4:00 271.9 F 29.2

Min 271.2 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 220.1 F 0.0
10:00 227.8 F 0.0
15:00 225.0 F 0.0
20:00 231.9 F 0.0
25:00 233.2 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:43
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

11/13/2023 11:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____