
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 003688
8:56 26/05/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:23
STERILIZING 3:17
136.0 C 220kPa 3:17
136.3 C 223kPa 3:47
136.4 C 223kPa 4:17
136.0 C 220kPa 4:47
136.4 C 223kPa 5:17
136.5 C 224kPa 5:47
136.4 C 221kPa 6:17
136.1 C 219kPa 6:47
136.1 C 219kPa 6:47
VENTING 6:47
AIR DRYING 7:33
CYCLE COMPLETE 7:38

Digital Signature #
D69B90E10671810D
