
Midmark M11 - v1.0.0
Total Cycles: 3271

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.7 F 0.0
2:00 198.0 F 0.3
4:00 209.9 F 5.4
6:00 220.0 F 10.6
8:00 225.2 F 6.1
10:00 237.7 F 11.5
12:00 236.6 F 9.1
14:00 239.0 F 9.4
16:00 249.2 F 14.7
18:00 261.5 F 21.5
20:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.9
2:00 271.8 F 28.1
3:00 272.1 F 28.7
4:00 272.6 F 28.7

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 218.7 F 0.0
10:00 235.8 F 0.0
15:00 232.5 F 0.0
20:00 222.0 F 0.0
25:00 218.0 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

05/23/2023 04:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____