
Midmark M11 - v1.0.0
Total Cycles: 1922

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2022 07:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 74.4 F 0.0
4:00 92.4 F 0.0
6:00 125.2 F 0.2
8:00 153.0 F 1.8
10:00 176.7 F 4.1
12:00 199.2 F 7.0
14:00 216.7 F 10.4
16:00 227.6 F 8.1
18:00 229.2 F 7.4
20:00 239.3 F 11.5
22:00 234.1 F 7.7
24:00 246.4 F 13.3
26:00 242.9 F 11.3
28:00 254.6 F 17.3
30:00 265.1 F 23.5
32:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.5 F 28.5
2:00 272.4 F 28.8
3:00 271.9 F 28.1
4:00 272.5 F 28.8

Min 271.6 F 28.1
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.4 F 0.0
5:00 205.3 F 0.0
10:00 223.7 F 0.0
15:00 201.3 F 0.0
20:00 200.0 F 0.0
25:00 201.8 F 0.0
30:00 206.6 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:40
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:10:27
CYCLE COMPLETE

05/10/2022 08:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____