
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002309
11:20 20/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:23
PRESSURIZING 1:30
STERILIZING 2:34
136.4 C 225kPa 2:34
136.1 C 224kPa 3:04
136.4 C 226kPa 3:34
136.4 C 226kPa 4:04
136.4 C 223kPa 4:34
136.2 C 224kPa 5:04
136.3 C 225kPa 5:34
136.4 C 223kPa 6:04
136.4 C 223kPa 6:04
VENTING 6:05
AIR DRYING 6:56
CYCLE COMPLETE 36:56

Digital Signature #
3B8E595704AE0D2D
