

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008475  
11:03 27/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:24  
STERILIZING 3:29  
136.5 C 223kPa 3:29  
136.4 C 225kPa 3:59  
136.4 C 220kPa 4:29  
136.4 C 223kPa 4:59  
136.4 C 224kPa 5:29  
136.4 C 225kPa 5:59  
136.4 C 223kPa 6:29  
136.4 C 224kPa 6:59  
136.4 C 224kPa 6:59  
VENTING 6:59  
AIR DRYING 7:49  
CYCLE COMPLETE 16:47

-----  
Digital Signature #  
CE9390C61C9F5F81  
-----