

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000984  
12:09 10/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:35  
STERILIZING 3:46  
136.4 C 232kPa 3:46  
136.3 C 233kPa 4:16  
136.4 C 233kPa 4:46  
136.4 C 233kPa 5:16  
136.2 C 231kPa 5:46  
136.4 C 232kPa 6:16  
136.4 C 233kPa 6:46  
136.4 C 233kPa 7:16  
136.4 C 233kPa 7:16  
VENTING 7:16  
AIR DRYING 8:07  
CYCLE COMPLETE 8:12  
-----

Digital Signature #  
89BB6EF4D38E4322  
-----