
Midmark M11 - v1.0.0
Total Cycles: 4547

Name: RITTER 1
Sterilizer ID: V2178198
05/13/2024 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.9 F 0.0
2:00 88.8 F 0.0
4:00 105.8 F 0.8
6:00 133.1 F 3.2
8:00 161.5 F 6.1
10:00 182.7 F 9.2
12:00 200.2 F 12.9
14:00 211.5 F 5.4
16:00 221.0 F 9.5
18:00 232.1 F 11.4
20:00 235.7 F 10.3
22:00 235.2 F 8.4
24:00 246.5 F 13.9
26:00 258.1 F 20.0
28:00 268.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.6
1:00 272.5 F 29.4
2:00 271.6 F 28.7
3:00 271.5 F 29.0
4:00 272.5 F 29.3

Min 271.0 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 216.0 F 0.0
10:00 214.1 F 0.0
15:00 223.2 F 0.0
20:00 229.1 F 0.0
25:00 232.9 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:19
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:48
CYCLE COMPLETE

05/13/2024 02:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____