
Midmark M11 - v1.0.0
Total Cycles: 994

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2021 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.6 F 0.0
2:00 202.1 F 2.9
4:00 199.5 F 7.5
6:00 211.6 F 12.7
8:00 229.4 F 8.5
10:00 241.8 F 11.6
12:00 242.2 F 11.6
14:00 240.2 F 10.2
16:00 253.8 F 17.1
18:00 265.1 F 23.8
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.1 F 28.9
2:00 272.4 F 29.1
3:00 272.5 F 29.1
4:00 272.4 F 29.1

Min 271.7 F 28.5
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 236.3 F 0.0
10:00 234.6 F 0.0
15:00 225.8 F 0.0
20:00 225.6 F 0.0
25:00 229.2 F 0.0
30:00 189.2 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

07/12/2021 11:09 AM
ACCEPTED: X
REJECTED:
Operator: U01