

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1273

Name: RITTER2  
Sterilizer ID: V2313228  
07/08/2022 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.0 F 0.0  
2:00 125.9 F 1.2  
4:00 141.6 F 4.2  
6:00 163.4 F 8.0  
8:00 188.2 F 12.8  
10:00 221.3 F 6.1  
12:00 228.4 F 10.8  
14:00 231.2 F 7.1  
16:00 242.2 F 12.4  
18:00 239.9 F 10.7  
20:00 252.4 F 17.0  
22:00 263.3 F 23.4  
24:00 272.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 273.0 F 30.2  
2:00 271.9 F 29.4  
3:00 272.4 F 30.1  
4:00 273.3 F 29.8

Min 270.9 F 28.9  
Max 273.5 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.7 F 0.0  
5:00 219.4 F 0.0  
10:00 217.5 F 0.0  
15:00 213.1 F 0.0  
20:00 203.9 F 0.0  
25:00 204.5 F 0.0  
30:00 184.6 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 24:15  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:02:42  
CYCLE COMPLETE

07/08/2022 11:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_