
Midmark M11 - v1.0.0
Total Cycles: 686

Name: MIDMARK
Sterilizer ID: V2178198
02/22/2021 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.9 F 0.0
2:00 126.8 F 0.8
4:00 188.6 F 3.1
6:00 199.8 F 7.8
8:00 213.1 F 13.1
10:00 228.2 F 6.9
12:00 235.5 F 12.2
14:00 237.0 F 9.2
16:00 235.9 F 9.3
18:00 247.8 F 14.2
20:00 260.5 F 21.2
22:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 271.9 F 28.8
2:00 271.8 F 28.8
3:00 271.9 F 28.7
4:00 272.3 F 28.7

Min 271.8 F 28.6
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 253.2 F 0.0
10:00 234.4 F 0.0
15:00 227.4 F 0.0
20:00 220.9 F 0.0
25:00 216.7 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:03
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

02/22/2021 12:56 PM
ACCEPTED: X
REJECTED:
Operator: U01