

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2628

Name: RITTER 1  
Sterilizer ID: V2178198  
11/30/2022 11:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.3 F 0.0  
2:00 189.4 F 4.5  
4:00 203.9 F 9.4  
6:00 219.9 F 6.9  
8:00 230.0 F 9.9  
10:00 232.7 F 8.2  
12:00 244.2 F 13.5  
14:00 243.9 F 12.5  
16:00 257.0 F 19.4  
18:00 268.6 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.6 F 29.3  
2:00 271.6 F 28.8  
3:00 272.7 F 29.4  
4:00 271.9 F 28.8  
  
Min 271.6 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 230.1 F 0.0  
10:00 230.0 F 0.0  
15:00 229.5 F 0.0  
20:00 227.7 F 0.0  
25:00 227.4 F 0.0  
30:00 227.0 F 0.0  
  
Drying Complete

FILLING: 1:41  
HEATING: 19:08  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:02  
CYCLE COMPLETE

11/30/2022 12:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_