
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002349
14:17 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:53
STERILIZING 3:57
136.4 C 226kPa 3:57
136.4 C 226kPa 4:27
136.3 C 225kPa 4:57
136.4 C 226kPa 5:27
136.4 C 225kPa 5:57
136.4 C 226kPa 6:27
136.3 C 223kPa 6:57
136.4 C 225kPa 7:27
136.4 C 225kPa 7:27
VENTING 7:27
AIR DRYING 8:22
CYCLE COMPLETE 35:59

Digital Signature #
AFDE3AB113BA5433
