

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2452

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2023 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
2:00 193.9 F 3.6  
4:00 203.2 F 8.5  
6:00 220.6 F 12.3  
8:00 229.5 F 8.7  
10:00 240.7 F 11.6  
12:00 241.4 F 11.9  
14:00 240.9 F 11.0  
16:00 253.9 F 17.6  
18:00 265.3 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.5 F 29.4  
2:00 271.5 F 28.6  
3:00 272.6 F 29.6  
4:00 272.1 F 29.0  
  
Min 271.4 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
5:00 217.7 F 0.0  
10:00 224.6 F 0.0  
15:00 222.2 F 0.0  
20:00 221.9 F 0.0  
25:00 217.8 F 0.0  
30:00 217.1 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 00:59:14  
CYCLE COMPLETE

06/01/2023 12:37 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_