
Midmark M11 - v1.0.5
Total Cycles: 4371

Name: RITTER2
Sterilizer ID: V2313228
02/21/2025 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.4 F 0.0
2:00 102.9 F 0.8
4:00 131.5 F 2.7
6:00 155.1 F 6.0
8:00 172.3 F 9.6
10:00 195.9 F 13.5
12:00 220.9 F 5.7
14:00 225.8 F 9.5
16:00 237.3 F 11.6
18:00 237.8 F 10.3
20:00 236.8 F 8.9
22:00 246.4 F 13.9
24:00 257.2 F 19.9
26:00 267.6 F 26.2
28:00 272.9 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.2 F 30.2
2:00 272.9 F 30.9
3:00 273.7 F 30.7
4:00 273.3 F 30.4

Min 271.0 F 29.3
Max 274.2 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 226.2 F 0.0
10:00 225.0 F 0.0
15:00 214.0 F 0.0
20:00 217.2 F 0.0
25:00 201.9 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 28:28
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:07:17
CYCLE COMPLETE

02/21/2025 04:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____