
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008242
11:24 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:01
STERILIZING 4:10
136.4 C 223kPa 4:10
136.1 C 221kPa 4:40
136.2 C 222kPa 5:10
136.4 C 224kPa 5:40
136.3 C 224kPa 6:10
136.3 C 223kPa 6:40
136.4 C 224kPa 7:10
136.4 C 224kPa 7:40
136.4 C 224kPa 7:40
VENTING 7:40
AIR DRYING 8:32
CYCLE COMPLETE 8:51

Digital Signature #
3B44017765C7462D
