
Midmark M11 - v1.0.0
Total Cycles: 4779

Name: RITTER 1
Sterilizer ID: V2178198
07/31/2024 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.7 F 0.0
2:00 96.8 F 1.1
4:00 129.8 F 3.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 137F
CHAMBER PRESS 4.4PSI
** ITEMS NOT STERILE **

07/31/2024 03:34 PM