
Midmark M11 - v1.0.0
Total Cycles: 4291

Name: RITTER 1
Sterilizer ID: V2178198
02/12/2024 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 88.6 F 0.2
4:00 86.5 F 1.2
6:00 116.8 F 3.2
8:00 142.1 F 6.0
10:00 165.1 F 9.5
12:00 186.9 F 13.0
14:00 211.4 F 3.4
16:00 216.5 F 6.4
18:00 226.6 F 9.7
20:00 233.9 F 13.5
22:00 235.3 F 9.1
24:00 241.8 F 11.2
26:00 243.8 F 12.1
28:00 255.5 F 18.1
30:00 265.8 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.3 F 28.8
2:00 271.8 F 28.3
3:00 271.9 F 28.6
4:00 272.2 F 28.6

Min 271.3 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.0 F 0.0
5:00 234.6 F 0.0
10:00 231.0 F 0.0
15:00 220.8 F 0.0
20:00 230.6 F 0.0
25:00 221.3 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:09:58
CYCLE COMPLETE

02/12/2024 10:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____