
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000738
10:39 30/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:40
STERILIZING 3:57
136.4 C 220kPa 3:57
136.1 C 220kPa 4:27
136.3 C 224kPa 4:57
136.4 C 221kPa 5:27
136.4 C 224kPa 5:57
136.4 C 218kPa 6:27
136.3 C 224kPa 6:57
136.4 C 221kPa 7:27
136.4 C 221kPa 7:27
VENTING 7:27
AIR DRYING 8:21
CYCLE COMPLETE 8:26

Digital Signature #
2FC2AACB19BB0329
