
Midmark M11 - v1.0.0
Total Cycles: 5494

Name:
Sterilizer ID: V2178198
04/10/2025 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.6 F 0.0
2:00 166.8 F 2.6
4:00 169.1 F 6.2
6:00 189.5 F 10.2
8:00 216.7 F 5.2
10:00 219.9 F 6.7
12:00 230.9 F 10.9
14:00 232.0 F 6.8
16:00 243.6 F 12.2
18:00 240.2 F 10.3
20:00 252.2 F 16.3
22:00 263.0 F 22.5
24:00 272.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.3 F 29.0
2:00 271.3 F 28.4
3:00 272.3 F 29.1
4:00 271.7 F 28.5

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 209.6 F 0.0
10:00 192.5 F 0.0
15:00 242.2 F 0.0
20:00 241.5 F 0.0
25:00 230.7 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:45
CYCLE COMPLETE

04/10/2025 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____