
Midmark M11 - v1.0.0
Total Cycles: 3656

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2023 01:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.3 F 0.0
2:00 160.3 F 3.3
4:00 179.4 F 7.1
6:00 208.3 F 12.4
8:00 219.3 F 6.5
10:00 232.2 F 11.6
12:00 232.7 F 8.2
14:00 242.3 F 11.7
16:00 244.4 F 12.6
18:00 256.6 F 19.0
20:00 267.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 271.9 F 28.5
2:00 272.3 F 28.8
3:00 272.5 F 29.2
4:00 271.8 F 28.7

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 218.7 F 0.0
10:00 232.9 F 0.0
15:00 228.1 F 0.0
20:00 228.0 F 0.0
25:00 227.2 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

08/15/2023 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____