

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2470

Name: RITTER 1  
Sterilizer ID: V2178198  
10/21/2022 12:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.6 F 0.0  
2:00 149.0 F 2.5  
4:00 172.9 F 5.9  
6:00 195.7 F 9.9  
8:00 211.6 F 4.1  
10:00 222.5 F 8.1  
12:00 235.4 F 13.1  
14:00 236.7 F 9.8  
16:00 238.6 F 8.6  
18:00 248.0 F 14.4  
20:00 259.9 F 20.9  
22:00 270.8 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.3 F 29.2  
2:00 271.8 F 28.5  
3:00 271.6 F 28.8  
4:00 272.2 F 29.1

Min 271.2 F 28.3  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 217.5 F 0.0  
10:00 206.7 F 0.0  
15:00 211.0 F 0.0  
20:00 242.0 F 0.0  
25:00 239.2 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 22:50  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:57  
CYCLE COMPLETE

10/21/2022 01:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_