

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003627  
11:02 09/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:40  
PRESSURIZING 3:40  
STERILIZING 4:46  
136.4 C 225kPa 4:46  
136.3 C 224kPa 5:16  
136.3 C 224kPa 5:46  
136.4 C 225kPa 6:16  
136.3 C 223kPa 6:46  
136.1 C 220kPa 7:16  
136.4 C 224kPa 7:46  
136.4 C 225kPa 8:16  
136.4 C 225kPa 8:16  
VENTING 8:16  
AIR DRYING 9:13  
CYCLE COMPLETE 9:19

-----  
Digital Signature #  
8222A06BEC8114CF  
-----