
Midmark M11 - v1.0.5
Total Cycles: 3433

Name: RITTER2
Sterilizer ID: V2313228
03/05/2024 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 72.1 F 0.0
4:00 83.5 F 0.0
6:00 107.2 F 1.1
8:00 143.1 F 3.7
10:00 168.9 F 6.7
12:00 185.2 F 10.1
14:00 205.3 F 5.8
16:00 213.7 F 6.4
18:00 226.6 F 10.4
20:00 231.3 F 8.4
22:00 239.6 F 11.7
24:00 237.7 F 9.8
26:00 250.6 F 16.0
28:00 261.4 F 22.2
30:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.4 F 29.5
2:00 271.4 F 29.0
3:00 272.6 F 29.5
4:00 271.3 F 28.9

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 209.8 F 0.0
10:00 196.3 F 0.0
15:00 192.9 F 0.0
20:00 188.7 F 0.0
25:00 198.0 F 0.0
30:00 204.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

03/05/2024 10:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____