
Midmark M11 - v1.0.0
Total Cycles: 5482

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2025 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 72.9 F 0.3
4:00 113.3 F 1.7
6:00 127.9 F 4.2
8:00 150.8 F 7.0
10:00 168.4 F 10.2
12:00 186.1 F 9.9
14:00 202.5 F 4.0
16:00 214.1 F 7.2
18:00 224.8 F 10.7
20:00 226.6 F 7.2
22:00 236.3 F 10.2
24:00 236.4 F 8.5
26:00 246.0 F 13.4
28:00 257.2 F 19.3
30:00 267.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.1
2:00 271.7 F 28.5
3:00 272.8 F 29.1
4:00 271.6 F 28.4

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 214.1 F 0.0
10:00 187.5 F 0.0
15:00 182.2 F 0.0
20:00 191.9 F 0.0
25:00 194.2 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

04/07/2025 09:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____