
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006037
7:44 25/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:51
STERILIZING 5:03
136.4 C 221kPa 5:03
136.4 C 224kPa 5:33
136.4 C 223kPa 6:03
136.4 C 221kPa 6:33
136.2 C 222kPa 7:03
136.1 C 221kPa 7:33
136.4 C 224kPa 8:03
136.4 C 221kPa 8:33
136.4 C 221kPa 8:33
VENTING 8:33
AIR DRYING 9:26
CYCLE COMPLETE 15:58

Digital Signature #
9AB83EA9AF203E95
