
Midmark M11 - v1.0.5
Total Cycles: 1883

Name: RITTER2
Sterilizer ID: V2313228
12/22/2022 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 182.1 F 3.3
4:00 191.9 F 8.2
6:00 211.2 F 10.8
8:00 226.0 F 9.0
10:00 235.4 F 9.9
12:00 242.7 F 13.0
14:00 243.1 F 12.3
16:00 256.3 F 19.2
18:00 267.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.9 F 29.6
2:00 271.7 F 28.9
3:00 272.9 F 29.9
4:00 272.5 F 29.3

Min 271.7 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 203.2 F 0.0
10:00 214.6 F 0.0
15:00 216.8 F 0.0
20:00 215.6 F 0.0
25:00 215.8 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 2:23
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

12/22/2022 03:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____