

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001049  
11:04 29/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:48  
STERILIZING 3:52  
136.0 C 221kPa 3:52  
136.3 C 224kPa 4:22  
136.3 C 224kPa 4:52  
136.0 C 221kPa 5:22  
136.4 C 222kPa 5:52  
136.4 C 225kPa 6:22  
136.3 C 224kPa 6:52  
136.3 C 223kPa 7:22  
136.3 C 223kPa 7:22  
VENTING 7:23  
AIR DRYING 8:12  
CYCLE COMPLETE 23:12  
-----

Digital Signature #  
D9451FC54C3BA2B6  
-----