

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000766  
10:31 28/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:12  
STERILIZING 4:21  
136.0 C 228kPa 4:21  
136.4 C 233kPa 4:51  
136.4 C 233kPa 5:21  
136.4 C 233kPa 5:51  
136.4 C 233kPa 6:21  
136.4 C 232kPa 6:51  
136.1 C 231kPa 7:21  
136.3 C 233kPa 7:51  
136.3 C 233kPa 7:51  
VENTING 7:51  
AIR DRYING 8:46  
CYCLE COMPLETE 8:50  
-----

Digital Signature #  
1F918B3708361393  
-----