
Midmark M11 - v1.0.0
Total Cycles: 4588

Name: RITTER 1
Sterilizer ID: V2178198
05/28/2024 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 154.2 F 2.0
4:00 160.5 F 5.1
6:00 185.4 F 8.5
8:00 200.1 F 12.8
10:00 212.6 F 5.3
12:00 223.9 F 9.5
14:00 235.0 F 10.8
16:00 237.5 F 10.5
18:00 235.7 F 8.3
20:00 247.9 F 14.5
22:00 259.0 F 20.5
24:00 269.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 29.3
2:00 272.0 F 28.8
3:00 271.2 F 28.6
4:00 272.3 F 29.4

Min 271.2 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 212.5 F 0.0
10:00 227.6 F 0.0
15:00 240.3 F 0.0
20:00 225.8 F 0.0
25:00 227.9 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:07
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:50
CYCLE COMPLETE

05/28/2024 10:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____