
Midmark M11 - v1.0.0
Total Cycles: 4791

Name: RITTER 1
Sterilizer ID: V2178198
08/07/2024 04:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.2 F 0.0
2:00 164.3 F 2.3
4:00 186.0 F 5.9
6:00 206.4 F 10.0
8:00 216.9 F 7.7
10:00 227.3 F 7.4
12:00 238.7 F 11.9
14:00 235.4 F 8.8
16:00 241.8 F 11.0
18:00 246.0 F 13.0
20:00 257.5 F 19.2
22:00 267.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.9
2:00 272.1 F 28.3
3:00 271.3 F 28.1
4:00 272.3 F 28.9

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 214.8 F 0.0
10:00 207.4 F 0.0
15:00 212.9 F 0.0
20:00 186.8 F 0.0
25:00 187.0 F 0.0
30:00 186.3 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

08/07/2024 05:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____