
Midmark M11 - v1.0.5
Total Cycles: 9

Name: RITTER2
Sterilizer ID: V2313228
04/22/2021 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.3 F 0.0
2:00 192.9 F 3.3
4:00 199.9 F 7.8
6:00 216.3 F 12.8
8:00 223.2 F 8.3
10:00 235.4 F 13.3
12:00 238.1 F 11.7
14:00 240.5 F 11.1
16:00 252.3 F 17.4
18:00 263.7 F 23.9
20:00 271.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 271.7 F 29.3
2:00 272.0 F 29.7
3:00 272.7 F 30.1
4:00 271.7 F 29.6

Min 271.0 F 29.1
Max 273.7 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 240.5 F 0.0
10:00 228.5 F 0.0
15:00 219.4 F 0.0
20:00 226.0 F 0.0
25:00 229.2 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 2:23
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:00
CYCLE COMPLETE

04/22/2021 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____