

-----  
Midmark M11 - v1.0.5  
Total Cycles: 742

Name: RITTER2  
Sterilizer ID: V2313228  
02/01/2022 09:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 101.0 F 0.0  
2:00 104.4 F 0.0  
4:00 117.8 F 0.1  
6:00 154.3 F 2.5  
8:00 183.3 F 6.4  
10:00 201.4 F 10.8  
12:00 218.6 F 5.3  
14:00 230.3 F 10.2  
16:00 232.5 F 8.1  
18:00 243.7 F 13.2  
20:00 242.9 F 11.9  
22:00 255.7 F 18.5  
24:00 267.0 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.1 F 28.9  
2:00 272.1 F 29.3  
3:00 271.7 F 28.6  
4:00 272.2 F 29.3

Min 271.3 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.3 F 0.0  
5:00 227.6 F 0.0  
10:00 208.4 F 0.0  
15:00 194.9 F 0.0  
20:00 187.5 F 0.0  
25:00 205.4 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 25:26  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:03:23  
CYCLE COMPLETE

02/01/2022 10:40 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_