

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2720

Name: RITTER2  
Sterilizer ID: V2313228  
08/01/2023 02:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
2:00 183.2 F 1.8  
4:00 192.6 F 7.1  
6:00 212.9 F 12.4  
8:00 225.2 F 7.8  
10:00 238.3 F 13.5  
12:00 239.8 F 11.6  
14:00 239.7 F 10.7  
16:00 253.7 F 17.6  
18:00 265.1 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.3 F 29.3  
2:00 272.6 F 29.8  
3:00 271.9 F 29.0  
4:00 272.7 F 29.8  
  
Min 271.8 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.5 F 0.0  
5:00 211.9 F 0.0  
10:00 228.4 F 0.0  
15:00 229.8 F 0.0  
20:00 227.3 F 0.0  
25:00 229.2 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:44  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

08/01/2023 03:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_