

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3204

Name: RITTER2  
Sterilizer ID: V2313228  
12/11/2023 01:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.2 F 0.0  
2:00 123.8 F 0.7  
4:00 147.4 F 3.3  
6:00 177.1 F 7.0  
8:00 197.4 F 11.0  
10:00 222.9 F 5.0  
12:00 226.8 F 9.6  
14:00 236.5 F 9.7  
16:00 241.4 F 11.7  
18:00 239.2 F 10.3  
20:00 252.3 F 16.8  
22:00 263.5 F 23.5  
24:00 271.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.8 F 30.0  
2:00 272.0 F 30.0  
3:00 271.4 F 29.6  
4:00 271.7 F 29.5

Min 270.4 F 28.7  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 169.9 F 0.0  
5:00 219.9 F 0.0  
10:00 188.4 F 0.0  
15:00 188.0 F 0.0  
20:00 187.0 F 0.0  
25:00 185.3 F 0.0  
30:00 183.7 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 24:04  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:02:39  
CYCLE COMPLETE

12/11/2023 02:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_