
Midmark M11 - v1.0.0
Total Cycles: 2203

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2022 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.4 F 0.0
2:00 202.2 F 4.5
4:00 208.5 F 9.6
6:00 221.4 F 6.8
8:00 231.3 F 10.0
10:00 235.6 F 7.9
12:00 243.8 F 12.7
14:00 241.6 F 11.1
16:00 254.6 F 17.7
18:00 266.1 F 24.6
20:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.9 F 29.5
2:00 271.9 F 28.8
3:00 273.0 F 29.6
4:00 272.3 F 29.0

Min 271.8 F 28.8
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 219.6 F 0.0
10:00 237.5 F 0.0
15:00 230.0 F 0.0
20:00 231.8 F 0.0
25:00 227.0 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

08/04/2022 04:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____