

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002430  
11:32 17/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:04  
STERILIZING 3:06  
136.0 C 222kPa 3:06  
136.4 C 226kPa 3:36  
136.2 C 224kPa 4:06  
136.2 C 224kPa 4:36  
136.3 C 225kPa 5:06  
136.2 C 225kPa 5:36  
136.4 C 226kPa 6:06  
136.0 C 225kPa 6:36  
136.0 C 225kPa 6:36  
VENTING 6:36  
AIR DRYING 7:25  
CYCLE COMPLETE 17:28  
-----

Digital Signature #  
C664AFD387C31DE6  
-----