
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001282
15:41 25/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:02
STERILIZING 3:48
136.0 C 220kPa 3:48
136.1 C 223kPa 4:19
136.3 C 224kPa 4:49
136.4 C 225kPa 5:19
136.4 C 221kPa 5:49
135.9 C 221kPa 6:19
136.3 C 224kPa 6:49
136.2 C 219kPa 7:19
136.2 C 219kPa 7:19
VENTING 7:19
AIR DRYING 8:13
CYCLE COMPLETE 8:19

Digital Signature #
E818EFD520F7CB37
