
Midmark M11 - v1.0.0
Total Cycles: 1329

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2021 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.8 F 0.0
2:00 121.6 F 0.0
4:00 134.8 F 0.9
6:00 163.5 F 3.5
8:00 193.5 F 7.6
10:00 216.7 F 12.1
12:00 222.3 F 6.2
14:00 233.9 F 10.8
16:00 236.2 F 7.2
18:00 242.5 F 12.6
20:00 240.3 F 10.6
22:00 252.5 F 16.6
24:00 262.9 F 22.7
26:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 273.2 F 29.2
2:00 271.9 F 28.4
3:00 272.2 F 29.0
4:00 272.4 F 29.0

Min 271.3 F 28.3
Max 273.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 224.9 F 0.0
10:00 221.7 F 0.0
15:00 218.7 F 0.0
20:00 226.4 F 0.0
25:00 216.5 F 0.0
30:00 215.3 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 26:22
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:04:40
CYCLE COMPLETE

11/02/2021 09:34 AM
ACCEPTED: X
REJECTED:
Operator: U01