
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000046
9:13 28/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 3:00
STERILIZING 4:32
136.0 C 221kPa 4:32
136.4 C 225kPa 5:02
136.4 C 225kPa 5:32
136.3 C 224kPa 6:02
136.0 C 222kPa 6:32
136.2 C 224kPa 7:02
136.4 C 223kPa 7:32
136.3 C 224kPa 8:02
136.3 C 224kPa 8:02
VENTING 8:02
AIR DRYING 8:55
CYCLE COMPLETE 21:19

Digital Signature #
C0F6B14F265C18DD
